

# GAGE-IT, INC.

## Pressure & Temperature Instrumentation

94 N. Branch Street  
Sellersville, PA 18960  
www.gageitinc.com

800-869-7294

(215) 453-8611  
Fax: (215) 453-8770  
e-mail: gageit@gageitinc.com

### CERTIFICATION OF ACCURACY FROM GAGE-IT, INC STANDARDS LABORATORY

Certification No: 132395

**SUBMITTED BY:**

Customer: Bellinger's Prof. Grounds Maint., Inc  
4902 State Rd 25 W  
Lafayette, IN 47909

Attn: Joshua Lane  
P.O. #: Verbal

Manufacturer: Backflow Test Equipment 845-3  
Description: Back Flow Test Kit  
Serial No: 12090127  
Customer Ref:  
Range: 0 - 15 P.S.I.D.  
Accuracy: +/- .2 P.S.I.D. DESC  
Pre-Cal: IN-SPEC Final: IN-SPEC

### STANDARD TRACEABILITY REPORT

Unit under test (UUT) conforms to present requirements of MIL-STD 45662A and ANSI/NC SL Z540-1-1994, PART II. Calibration is traceable to N.I.S.T. and is within manufacturers specifications, unless noted.

### STANDARDS USED

<u>Description</u>	<u>N.I.S.T. Traceability</u>	<u>Standard Due Date</u>
Additel ADT681 Digital Gauge, 0-100#	N19092707	9/26/2020

### CALIBRATION DATA

Standard P.S.I.D.	UUT As Rcvd P.S.I.D.	UUT As Left P.S.I.D.	Min	Max
7.0	7.0	7.0	6.8	7.2
5.0	5.0	5.0	4.8	5.2
2.0	2.0	2.0	1.8	2.2
1.0	1.0	1.0	0.8	1.2

Temperature: 68 Degrees F  
Humidity: 25 % RH  
Void Seals: 0

Date Tested: 12/27/2019  
Date Due: 12/27/2020  
Certified By: Eric Stecker  
Technician  
Approved By: Amy Stecker  
Quality Control

